



## CERTIFICATE OF ANALYSIS

PO BOX 130549 Spring, TX 77393  
Phone: (281) 298-9410 Fax: (281) 298-9411

<b>FINISHED PRODUCT, LOT NUMBER, MFG. /EXP DATE:</b> PolySeed® • Part No. P-110 • Lot 212008 • Mfg. Date: 08/2020 • Exp. Date: 08/2022
<b>FORMULATION:</b> The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.
<b>VIABLE COUNT, FINAL TEST RESULT:</b> The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of $4.00 \times 10^9$ cfu/g
<b>GLUCOSE/GLUTAMIC-ACID RESULTS:</b> Tested results within acceptable range 198 +/- 30.5 mg/L (167.5 - 228.5 mg/L). GGA Lot# J172-23 – Average Test Result: 204.6  See <a href="http://www.polyseed.com">www.polyseed.com</a> for details
<b>SEED CONTROL FACTOR:</b> Tested results within acceptable range 0.6 – 1.0 see <a href="http://www.polyseed.com">www.polyseed.com</a> for details
<b>SALMONELLA TEST RESULT:</b> The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.
The purpose of this document is to assure that the Finished Product conforms to the above specification.  <p style="text-align: center;"> <b>Signature:</b> _____ <b>Date:</b> 08/21/2020 <i>Quality Control Department</i></p>
POLYSEED.Ref.1.19 <span style="float: right;">Revised Jan 20</span>