



CERTIFICATE OF ANALYSIS

PO BOX 130549 Spring, TX 77393
Phone: (281) 298-9410 Fax: (281) 298-9411

FINISHED PRODUCT, LOT NUMBER, MFG. /EXP DATE:

PolySeed® • Part No. P-110 • Lot 252103 • Mfg. Date: 03/2021 • Exp. Date: 03/2023

FORMULATION:

The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.

VIABLE COUNT, FINAL TEST RESULT:

The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of 4.00×10^9 cfu/g.

GLUCOSE/GLUTAMIC-ACID RESULTS:

Tested results within acceptable range 198 +/- 30.5 mg/L (167.5 - 228.5 mg/L). GGA Lot# J317-19 – Average Test Result: 210.8

See www.polyseed.com for details.

SEED CONTROL FACTOR:

Tested results within acceptable range 0.6 – 1.0 see www.polyseed.com for details

SALMONELLA TEST RESULT:

The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.

The purpose of this document is to assure that the Finished Product conforms to the above specification.

Signature: 

Quality Control Department

Date: 03/25/2021

POLYSEED.Ref.1.19

Revised Jan 21